



	Control Standard Recovery (%)	Values in µg/L	
		Bellville 1	Bellville 2
<b>B</b>	92.9	7.38	4.10
<b>Al</b>	102.2	38.43	40.34
<b>Ti</b>	107.4	0.66	0.48
<b>V</b>	98.7	0.08	0.13
<b>Cr</b>	99.4	0.25	0.25
<b>Mn</b>	100.0	3.83	3.40
<b>Fe</b>	98.0	18.80	19.34
<b>Co</b>	100.0	0.01	0.01
<b>Ni</b>	97.9	0.15	0.11
<b>Cu</b>	99.5	1.98	2.48
<b>Zn</b>	98.2	3.32	1.29
<b>As</b>	103.5	0.29	0.30
<b>Se</b>	93.7	0.50	0.36
<b>Mo</b>	100.2	0.92	0.57
<b>Cd</b>	96.3	<0.000	<0.000
<b>Sb</b>	92.5	0.01	0.01
<b>Ba</b>	101.3	1.58	0.98
<b>Hg</b>	105.3	0.01	0.01
<b>Pb</b>	97.5	0.08	0.04
		Values in mg/L	
		Bellville 1	Bellville 2
<b>Ca</b>	102.3	10.46	10.33
<b>K</b>	98.1	0.33	0.36
<b>Mg</b>	105.4	1.06	1.05
<b>Na</b>	101.9	5.35	5.37

Date completed: 15/12/2014

Comments

We are not an accredited laboratory, however International quality control guidelines and certified calibration standards are used to verify accuracy of results.

Analyst & ICP-MS Solutions Manager:

Riana Rossouw